

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 08/31/2015

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 15140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:38
INTERNAL STANDARD	VERIFIED	02:38
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:42

*SOS RJB* *Bartlett* *WCO #142*  
08-31-15

