

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/08/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:42
INTERNAL STANDARD	VERIFIED	03:42
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:42
BLANK TEST	0.00	03:43
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	03:43
BLANK TEST	0.00	03:44
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	03:45
BLANK TEST	0.00	03:46



SGT Rk J Bartlett WCSO #142
09-08-15