

SUPERVISOR TEST

WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/17/2015

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 15140

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 01:35 |
| INTERNAL STANDARD       | VERIFIED | 01:35 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.09     | 01:36 |
| BLANK TEST              | 0.00     | 01:37 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.09     | 01:37 |
| BLANK TEST              | 0.00     | 01:38 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.09     | 01:38 |
| BLANK TEST              | 0.00     | 01:39 |

SGS PLS 1 Runoff WOSO #162  
09-17-15

