

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/22/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:20
INTERNAL STANDARD	VERIFIED	02:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:20
BLANK TEST	0.00	02:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:22
BLANK TEST	0.00	02:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:24



565 *Ralph Bartlett* WCSO #162
09-22-15