

SUPERVISOR TEST




WARREN COUNTY  
 SHERRIF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 10/06/2015

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 15180

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:34
INTERNAL STANDARD	VERIFIED	01:34
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:34
BLANK TEST	0.00	01:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:36
BLANK TEST	0.00	01:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:37
BLANK TEST	0.00	01:38



*SG Bartlett* WESB #162  
 10-06-15