

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 10/12/2015

OPERATOR NAME:
 KENNETH W SMITH
 Permit #: 35604
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15180

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:56
INTERNAL STANDARD	VERIFIED	04:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:57
BLANK TEST	0.00	04:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:58
BLANK TEST	0.00	04:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:59
BLANK TEST	0.00	05:00



PO. Kenneth W. Smith # 301
 10/12/15