

SUPERVISOR TEST



WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 10/20/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15180

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:25
INTERNAL STANDARD	VERIFIED	02:25
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:25
BLANK TEST	0.00	02:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:26
BLANK TEST	0.00	02:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:27
BLANK TEST	0.00	02:28



SGS *RJ Bartlett* WESO #162
 10-20-15