

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/25/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15180

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:16
INTERNAL STANDARD	VERIFIED	04:16
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	04:16
BLANK TEST	0.00	04:17
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	04:18
BLANK TEST	0.00	04:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:19
BLANK TEST	0.00	04:20



Sgt. Ralph Bartlett WSO #162
10-25-15