

SUPERVISOR TEST




WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 12/09/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15210

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:55
INTERNAL STANDARD	VERIFIED	23:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:55
BLANK TEST	0.00	23:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:56
BLANK TEST	0.00	23:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:58
BLANK TEST	0.00	23:58



SUS RLY Thru was #162
 12/9/15