

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/14/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15210

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:28
INTERNAL STANDARD	VERIFIED	16:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	16:29
BLANK TEST	0.00	16:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	16:30
BLANK TEST	0.00	16:31
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	16:31
BLANK TEST	0.00	16:32



SGS RJP Bartlett waso #162
12-14-15