

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 12/29/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15210

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:54
INTERNAL STANDARD	VERIFIED	02:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:56
BLANK TEST	0.00	02:57
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:58



SK Rep / Bartlett wsn #102
 12-29-15