

SOLUTION CHANGE




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 05/03/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16050

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:23
INTERNAL STANDARD	VERIFIED	01:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:24
BLANK TEST	0.00	01:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:27
BLANK TEST	0.00	01:28



SGT *Ralph Bartlett* WCSO #162
 05-03-16