

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/18/2016

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:48
INTERNAL STANDARD	VERIFIED	02:49
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:49
BLANK TEST	0.00	02:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:50
BLANK TEST	0.00	02:51
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:52
BLANK TEST	0.00	02:53

P.O. Kenneth W. Smith
7/18/16

