

SUPERVISOR TEST

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 07/27/2016

OPERATOR NAME:
 KENNETH W SMITH
 Permit #: 35604
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:38
INTERNAL STANDARD	VERIFIED	02:38
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:39
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:41
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:42

PK [Signature]
 7/27/16

