

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 10/10/2016

OPERATOR NAME:  
JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 16170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:09
INTERNAL STANDARD	VERIFIED	09:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:09
BLANK TEST	0.00	09:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:11
BLANK TEST	0.00	09:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:12
BLANK TEST	0.00	09:13

*J.M. Lopez 299*

