

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 11/03/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 2622
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16170

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:51
INTERNAL STANDARD	VERIFIED	15:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:51
BLANK TEST	0.00	15:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:53
BLANK TEST	0.00	15:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:54
BLANK TEST	0.00	15:55



SG Bartlett WCSO #162
 11/3/16