

SUPERVISOR TEST

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 12/19/2016

OPERATOR NAME:
 KENNETH W SMITH
 Permit #: 35604
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:19
INTERNAL STANDARD	VERIFIED	01:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:21
BLANK TEST	0.00	01:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:22
BLANK TEST	0.00	01:23

K. W. Smith
 12/19/16

