

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/01/2016

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15250

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:35
INTERNAL STANDARD	VERIFIED	02:35
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:36
BLANK TEST	0.00	02:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:37
BLANK TEST	0.00	02:38
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:40



SGS Del 7 *[Signature]* was #162
02-01-16