

SOLUTION CHANGE




WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106606

Date: 04/04/2016

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 16050

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:40
INTERNAL STANDARD	VERIFIED	23:40
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.10	23:41
BLANK TEST	0.00	23:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:42
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:43
BLANK TEST	0.00	23:44



SCS *Ralph J Bartlett*  
 WCO #162 04-04-16