

SUPERVISOR TEST




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 04/19/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16050

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:25
INTERNAL STANDARD	VERIFIED	00:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:26
BLANK TEST	0.00	00:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:27
BLANK TEST	0.00	00:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:28
BLANK TEST	0.00	00:29



SOS Ralph Bartlett weso #162
 04-19-16