

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 06/01/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:49
INTERNAL STANDARD	VERIFIED	02:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:49
BLANK TEST	0.00	02:50
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:51
BLANK TEST	0.00	02:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:52
BLANK TEST	0.00	02:53



*Sis Kelly Bates was #162
 06/01/16*