

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/12/2016

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:32
INTERNAL STANDARD	VERIFIED	04:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:33
BLANK TEST	0.00	04:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:34
BLANK TEST	0.00	04:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:36
BLANK TEST	0.00	04:37

P.O. [Signature]

7/12/16

