

SUPERVISOR TEST

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 07/18/2016

OPERATOR NAME:
 KENNETH W SMITH
 Permit #: 35604
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:00
INTERNAL STANDARD	VERIFIED	04:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:01
BLANK TEST	0.00	04:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:02
BLANK TEST	0.00	04:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:03
BLANK TEST	0.00	04:04

PO. Kenneth W Smith #301

7/18/16

