

SUPERVISOR TEST

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 07/27/2016

OPERATOR NAME:
 KENNETH W SMITH
 Permit #: 35604
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:18
INTERNAL STANDARD	VERIFIED	04:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:19
BLANK TEST	0.00	04:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	04:20
BLANK TEST	0.00	04:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:21
BLANK TEST	0.00	04:22

AC [Signature] #301

7/27/16

