

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 08/01/2016

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16120

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:10
INTERNAL STANDARD	VERIFIED	00:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:11
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:13
BLANK TEST	0.00	00:14



SCS *RJ Bartlett*
weso #162 08-01-16