

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 08/09/2016

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16120

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:35
INTERNAL STANDARD	VERIFIED	02:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:35
BLANK TEST	0.00	02:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:37
BLANK TEST	0.00	02:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:38
BLANK TEST	0.00	02:39



*Sus Rly Results were #162
08-09-16*