

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 10/02/2016

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:15
INTERNAL STANDARD	VERIFIED	21:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:16
BLANK TEST	0.00	21:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:17
BLANK TEST	0.00	21:18
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	21:18
BLANK TEST	0.00	21:19



SCS Dept of Corrections Wesc#162
10-07-16