

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 10/10/2016

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16170

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:21
INTERNAL STANDARD	VERIFIED	10:21
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	10:21
BLANK TEST	0.00	10:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:22
BLANK TEST	0.00	10:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:24
BLANK TEST	0.00	10:24

P.C. gmt 29c

