

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 11/03/2016

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16170

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:17
INTERNAL STANDARD	VERIFIED	15:18
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	15:18
BLANK TEST	0.00	15:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:19
BLANK TEST	0.00	15:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:21
BLANK TEST	0.00	15:22



SGT R. J. Bartlett w/cso #162 11/3/16