

SOLUTION CHANGE




WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106606

Date: 11/29/2016

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:52
INTERNAL STANDARD	VERIFIED	02:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:54
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:56



SGT Dept of Justice WLSO #162  
 11/29/16