

SUPERVISOR TEST




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 12/13/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:59
INTERNAL STANDARD	VERIFIED	01:59
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:00
BLANK TEST	0.00	02:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:01
BLANK TEST	0.00	02:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:02
BLANK TEST	0.00	02:03



SR *Ralph Bartlett* Weso #162
 12-13-16