

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 12/19/2016

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:46
INTERNAL STANDARD	VERIFIED	01:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:46
BLANK TEST	0.00	01:47
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:48
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:49
BLANK TEST	0.00	01:50

PO [Signature] # 301

12/19/16

