

SUPERVISOR TEST




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 12/26/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:42
INTERNAL STANDARD	VERIFIED	23:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:42
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	23:43
BLANK TEST	0.00	23:44
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:44
BLANK TEST	0.00	23:45



SGT Ralph Bartlett WSO #162
 12-26-16