

SOLUTION CHANGE




 WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 01/04/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15210

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:36
INTERNAL STANDARD	VERIFIED	01:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:37
BLANK TEST	0.00	01:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:38
BLANK TEST	0.00	01:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:39
BLANK TEST	0.00	01:40



SGT *Ralph J Bartlett*
 WC SO # 162 01-04-16