

SUPERVISOR TEST




WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/26/2016

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15210

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:56
INTERNAL STANDARD	VERIFIED	03:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:58
BLANK TEST	0.00	03:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:59
BLANK TEST	0.00	04:00



SCR Ralph Bartlett WCSO #162
01-26-16