

SOLUTION CHANGE




WARREN COUNTY  
 SHERRIF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 02/01/2016

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 15250

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:54
INTERNAL STANDARD	VERIFIED	02:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:54
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:57



SGS *Ralph J Bartlett* *Permit #162*  
 02-01-16