

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/17/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15250

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:55
INTERNAL STANDARD	VERIFIED	12:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:55
BLANK TEST	0.00	12:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:56
BLANK TEST	0.00	12:57
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	12:58
BLANK TEST	0.00	12:59



SGP Ralph Bartlett WSO #162
 02-17-16