

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 03/08/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15250

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:06
INTERNAL STANDARD	VERIFIED	02:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:07
BLANK TEST	0.00	02:08
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:09
BLANK TEST	0.00	02:10



Sgt Ray Bartlett WESB #122
 03-08-16