

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 03/28/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15250

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:20
INTERNAL STANDARD	VERIFIED	00:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:20
BLANK TEST	0.00	00:21
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:22
BLANK TEST	0.00	00:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:23
BLANK TEST	0.00	00:24



SCS Ralph J Bartlett
 WCSO #162 03-28-16