

SOLUTION CHANGE



WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/05/2016

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16050

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:18
INTERNAL STANDARD	VERIFIED	00:19
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:19
BLANK TEST	0.00	00:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:20
BLANK TEST	0.00	00:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:21
BLANK TEST	0.00	00:22



SGS *[Signature]* *[Signature]* waso #162
04-05-16