

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/11/2016

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16050

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:28
INTERNAL STANDARD	VERIFIED	00:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:28
BLANK TEST	0.00	00:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:29
BLANK TEST	0.00	00:30
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:32



565 *Shelley* *Barrett* *work #162*
04-11-16