

SUPERVISOR TEST

WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 04/19/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16050

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:40
INTERNAL STANDARD	VERIFIED	00:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:43
BLANK TEST	0.00	00:44

505 Rt / Bartlett waso #162
 04-19-16

