

SUPERVISOR TEST

WARREN COUNTY  
 SHERRIF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 04/25/2016

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 16050

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:29
INTERNAL STANDARD	VERIFIED	00:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:30
BLANK TEST	0.00	00:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:32
BLANK TEST	0.00	00:33

*Sgt R J Bartlett waco #162*  
 04-25-16

