

SUPERVISOR TEST



WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/09/2016

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16050

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:24
INTERNAL STANDARD	VERIFIED	00:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:24
BLANK TEST	0.00	00:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:26
BLANK TEST	0.00	00:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:27
BLANK TEST	0.00	00:28



see Ralph J Bartlett WCSO#162
05-09-16