

SUPERVISOR TEST




WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/22/2016

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 16050

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:51
INTERNAL STANDARD	VERIFIED	22:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:51
BLANK TEST	0.00	22:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:52
BLANK TEST	0.00	22:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:54
BLANK TEST	0.00	22:55



SGT Rudy [Signature] Weso #162  
05-22-16