

SOLUTION CHANGE




WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/01/2016

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:13
INTERNAL STANDARD	VERIFIED	01:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:14
BLANK TEST	0.00	01:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:15
BLANK TEST	0.00	01:16
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:16
BLANK TEST	0.00	01:17



SGS Supp. Burey was #1/c-
06/01/16