

SUPERVISOR TEST



WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 06/06/2016

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:32
INTERNAL STANDARD	VERIFIED	02:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:32
BLANK TEST	0.00	02:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:33
BLANK TEST	0.00	02:34
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:34
BLANK TEST	0.00	02:35



565 Ref / Bartlett  
wco #162 06-06-16