

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/14/2016

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:43
INTERNAL STANDARD	VERIFIED	02:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:43
BLANK TEST	0.00	02:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:44
BLANK TEST	0.00	02:45
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:46
BLANK TEST	0.00	02:47



Sgt Ralph Bartlett WESO # 16090
06-14-16