

SUPERVISOR TEST




WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 06/28/2016

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:24
INTERNAL STANDARD	VERIFIED	01:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:24
BLANK TEST	0.00	01:25
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:26
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:27
BLANK TEST	0.00	01:28



SGS RL / Bartlett  
WCSO #162 06-28-16