

SUPERVISOR TEST

WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 07/12/2016

OPERATOR NAME:  
KENNETH W SMITH  
Permit #: 35604  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:51
INTERNAL STANDARD	VERIFIED	03:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:51
BLANK TEST	0.00	03:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:53
BLANK TEST	0.00	03:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:54
BLANK TEST	0.00	03:55

*P.O. [Signature] #301*

*7/12/16*

